

MEDICAL DEVICE CYBERSECURITY


SETH D. CARMODY, PHD

APRIL 16, 2018

AMDM 2018 IVD SUBMISSIONS WORKSHOP

Intended Use + Misuse

<http://hackaday.com/2015/09/07/brick-laying-robot-does-it-better/>



<http://www.technologyvista.in/pin/here-comes-the-brick-laying-robot-to-make-buildings/>

Negative Requirements are *Infinite*



The diagram consists of two overlapping circles. The top circle is purple and contains the text 'Features: What a Device MUST Do...'. The bottom circle is blue and contains the text 'Safety: What a Device MUST NOT do'. The intersection of the two circles is shaded a darker purple and contains the text 'Thou, shall not under or over deliver therapy!'. The entire diagram is enclosed within a larger, light orange circle with a dashed orange border.

Features:
What a Device
MUST Do...

Get drug libraries
from the Internet

Safety:
What a Device
MUST NOT do

Thou, shall not
under or over
deliver therapy!

Device Lifecycle: Ecosystem Challenges



Let's Play: Fact vs. Myth


Fact vs. Myth

The FDA is the federal entity solely responsible for the cybersecurity of medical devices.


Myth

FACT: The FDA works closely with several federal government agencies including the U.S. Department of Homeland Security (DHS), members of the private sector, medical device manufacturers, healthcare delivery organizations, security researchers, and end users to increase the security of the U.S. critical cyber infrastructure.

A Complex Ecosystem



FDA Cybersecurity History



Premarket Cybersecurity Guidance

- Draft June 2013
- Final October 2014
- Key Principles:
 - #1 Shared responsibility between stakeholders, including health care facilities, patients, providers, and manufacturers of medical devices
 - #2 Address cybersecurity during the design and development of the medical device
 - #3 Establish design inputs for device related to cybersecurity, and establish a cybersecurity vulnerability and management approach as part of the software validation and risk analysis that is required by 21 CFR 820.30(g)

Let's Play: Fact vs. Myth

Fact vs. Myth

Cybersecurity for medical devices is optional.

Myth

FACT: Medical device manufacturers must comply with federal regulations. Part of those regulations, called quality system regulations (QSRs), requires that medical device manufacturers address all risks, including cybersecurity risk. The pre- and post- market cybersecurity guidances provide recommendations for meeting QSRs.

Let's Play: Fact vs. Myth

Fact vs. Myth

The FDA does not conduct premarket testing for any medical products.

FACT

Myth: The FDA tests for cybersecurity of medical devices.

Key Principles of FDA Postmarket Management of Cybersecurity in Medical Devices



- Use a risk-based framework to assure risks to public health are addressed in a continual and timely fashion
- Articulate manufacturer responsibilities by leveraging existing Quality System Regulation and postmarket authorities
- Foster a collaborative and coordinated approach to information sharing and risk assessment
- Align with Presidential EOs and NIST Framework
- Incentivize the “right” behavior

Let's Play: Fact vs. Myth

Fact vs. Myth

Medical device manufacturers can't update medical devices for cybersecurity.

Myth

FACT: Medical device manufacturers can always update a medical device for cybersecurity. In fact, the FDA does not typically need to review changes made to medical devices solely to strengthen cybersecurity

Let's Play: Fact vs. Myth

Fact vs. Myth

Health care Delivery Organizations (HDOs) can't update and patch medical devices for cybersecurity.

Myth


FACT: The FDA recognizes that HDOs are responsible for implementing devices on their networks and may need to patch or change devices and/or supporting infrastructure to reduce security risks. Recognizing that changes require risk assessment, the FDA recommends working closely with medical device manufacturers to communicate changes that are necessary.

Postmarket Cybersecurity Risk Assessment



Severity of Patient Harm (if exploited)

Negligible Minor Serious Critical Catastrophic



Assessing Exploitability with Common Vulnerability Scoring System (CVSS)

- Establish a repeatable process by leveraging existing frameworks (e.g. CVSS)

Base Scoring (risk factors of the vulnerability)

e.g. Attack Vector (physical, local, adjacent, network)


Temporal Scoring (risk factors that change over time)

e.g. Exploit Code Maturity (high, functional, proof-of-concept, unproven)

Environmental scoring (controls that reduce risk)

e.g. Physical, software, network, compensating controls.

Changes to a Device for Controlled vs. Uncontrolled Risk



Questions?

Contacts:

CDRH mailbox, CyberMed@fda.hhs.gov

Suzanne Schwartz, Suzanne.Schwartz@fda.hhs.gov

Aftin Ross, aftin.ross@fda.hhs.gov

Seth Carmody, seth.carmody@fda.hhs.gov