

# Special and Abbreviated 510(k)s & Add-to files

Office of In Vitro Diagnostics  
Submissions Workshop  
April 25 – 26, 2011

Doug Rheinheimer  
Division of Chemistry and Toxicology Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety



# Topics

---

- What to do when modifying an existing device
- What is a special 510(k) and when can it be used?
  - What to submit for a special 510(k)
- What is an abbreviated 510(k)?
  - What to submit for an abbreviated 510(k)
- What are some advantages/disadvantages of each
- When to use add-to file submission

# Modifying an existing device

---

You have a device in commercial distribution but a modification may significantly affect safety or effectiveness.

Start with our Guidance – “Deciding When to Submit a 510(k) for a Change to an Existing Device” (K97-1)


Guidance includes a flowchart model that can help in your decision-making to analyze how changes in devices may affect safety or effectiveness and if a new 510(k) is required

# Flowcharts


---


- Flowchart A - labeling changes
- Flowchart B - technology or performance specifications changes
- Flowchart C - materials changes
- Flowchart D - materials changes for *in vitro* devices (IVDs)

## FLOWCHART A - IS IT A LABELING CHANGE?



# **FLOWCHART B - IS IT A TECHNOLOGY OR PERFORMANCE CHANGE?**





# **New 510(k)? 5 Questions from Flowchart D (materials change for IVDs)**

---

- D1 Change in performance specifications?  
(cutoff, expected values, precision,  
interferences)
- D1.1 Is new clinical data (clinical samples)  
necessary to establish safety and  
effectiveness?
- D.1.2 Is new clinical data necessary to show  
continuing conformance of the device to a  
recognized standard? (CRMLN, NGSP)



# New 510(k)? 5 Questions

---

- D.1.3 Results of the design validation performed as a result of change in materials raise new issues of safety and effectiveness?
- D2 Change in material alter the operating principle of the IVD?

# New 510(k)?

---

If all “no” answers:

- Any modifications must be made in accordance with the Quality System regulation, 21 CFR 820, and recorded in the device master record and change control records. This info should be available for FDA inspection.
- Keep justification for (not) submitting a new 510(k) in the change control records.
- Do not need to send “Add-to file” saying you made changes (more on this later)

# New 510(k)?


---

- If the flowchart leads you to “New 510(k)”, then the next question is: special, traditional or abbreviated?

There’s a guidance for that – “The New 510(k) Paradigm - Alternate Approaches to Demonstrating Substantial Equivalence in Premarket Notifications”

# The New 510(k) Paradigm

At



# What is a special 510(k)?

---

- Utilizes the design control requirement of the Quality System Regulation (21 CFR 820)
- Can only be submitted for a modification to your own device that has been cleared under the 510(k) process
- Allows the manufacturer to declare conformance to design controls without providing the raw data

# Design Control Requirements

---

- Require the manufacturer to conduct verification and validation studies of a type that traditionally may have been included in 510(k) submissions
- May be appropriate to forgo a detailed review of the underlying data normally required in a traditional 510(k).

# Three questions to address

---

- Change in intended use/indications for use or any labeling change that affects intended use?
  - Change in fundamental scientific technology?
  - Change requiring clinical study to evaluate patient safety and effectiveness?
- 
- If all are no = may be eligible for a special
  - If any are yes = traditional or abbreviated 510(k)

# Changes typically eligible for a special 510(k)

---

- Change in reagent (dry to liquid)
- Change in expiration dating
- Change in manufacturing to produce reagents that do not need calibration by user
- Adding an additional anticoagulant as an acceptable sample



# Not Eligible for a Special 510(k)

---

- Change in intended/indications for use
  - Change from prescription use to OTC
- Change in derivation of algorithm
- Change in major reactive ingredient that affects patient safety and effectiveness

# **Not Eligible for a Special 510(k)**

---

- Change in cut-off that needs a clinical study to assess patient safety and effectiveness
- Combining two cleared devices to make new

# What to submit for a special 510(k)

---

Administratively, there are 11 items:

- Medical Device User Fee Cover Sheet (Form FDA 3601)
- CDRH Premarket Review Submission Cover Sheet (Form FDA 3514)
- Certification of Compliance with ClinicalTrials.gov Data Bank (Form FDA 3674)

# What to submit for a special 510(k)

---

- Cover Letter, identifying the application as a "Special 510(k)." Include 510(k) holder name, address, and facility registration number, if available.
- Table of Contents
- 510(k) Screening Checklist
- Statement of Indications for Use (OIVD form)

# What to submit for a special 510(k)

---

- 510(k) Summary [21 CFR 807.92] or 510(k) Statement [21 CFR 807.93]
- Standards Data Report for 510(k)s – Form FDA 3654. Submit this form if your 510(k) references a national or international standard.
- Truthful and Accuracy Statement
- Declaration of Conformity

# What to submit for a special 510(k)

---

- Detailed description of modified device (including Ind. For Use) and *any and all* differences between modified and cleared device
- Provide a statement that there is no change in fundamental technology and no change in intended use
- Proposed labeling with all changes from predicate highlighted or prominently identified
- Summary of design control activities

# Summary of Design Control Activities

---



- State risk analysis method used to assess the impact of the modification
- Provide all verification/validations tests that were performed, *as required by the risk analysis*
- List pre-determined acceptance criteria appropriate for the clinical needs of assay and a summary of results showing pre-determined acceptance criteria were met
- Provide a statement that pre-determined acceptance criteria were met

# Signed Declaration of Conformity with design control requirements

---



- All verification activities, as required by the risk analysis, for the modification were performed by the designated individual(s) and the results demonstrated that the predetermined acceptance criteria were met
- The manufacturing facility, *[Company Name]* is in conformance with the design control requirements as specified in 21 CFR 820.30 and the records are available for review.



# **FDA Reviewer's Expectations for a Special 510(k)**

---

- Validation/verification activities similar to those in traditional 510(k): method comparison, linearity, precision, interference studies, etc.
- Graphs and/or charts of data analyses clearly showing acceptance criteria were met and verification/validation activities are complete

# Advantages of a Special 510(k)

---



- 30 days for decision by FDA
- Declare conformance to design control – should be easier to prepare

# Potential Problems with Special 510(k)s


---

Most misunderstood type of submission

- May not be clearly presented
- FDA does not understand the modifications
- Acceptance criteria not clinically relevant or are not met
- Risk analysis and/or verification & validation activities are not relevant to the modification

# The New 510(k) Paradigm

Attachment 1



This flowchart should only be considered in conjunction with the accompanying proposed text.

# What is an Abbreviated 510(k)?

---

Device manufacturers may choose to submit an Abbreviated 510(k) when:

- There is a device-specific guidance document
- A special control has been established
- FDA has recognized a relevant consensus standard

# Abbreviated 510(k)

---

Need a summary report -

- Describes adherence to the relevant guidance or special control and how they were used during device development and testing
- Declaration of conformity if using standard

# Challenges of Abbreviated 510(k)s

---

- One size does not fit all –
- Relatively few guidance documents to cover all aspects of IVD device performance
- No time advantage over traditional (90 FDA review days)

# Advantages of Abbreviated 510(k)s

---

- Suitable for submissions for calibrator or control materials
- Potentially easier to prepare



# What to Submit for an Abbreviated 510(k) – Guidance Documents

---



- All the administrative information required for a traditional 510(k)
- A summary report describing adherence to the relevant guidance document and how the document was used during device development and testing, including the manufacturer's efforts to conform with the guidance document and any deviations
- A summary report that describes how the guidance document was used to address the risks associated with the particular device type
- Data/information to address issues not covered by guidance documents, special controls, and/or recognized standards
- Information on sterilization, biocompatibility, expiration date, etc., if applicable.

# What to Submit for an Abbreviated 510(k) – Special Controls

---

- All the administrative information required for a traditional 510(k)
- A summary report that describes adherence to the special control and how the special control(s) was used during device development and testing, including to address a specific risk or issue with the device. The report should include the manufacturer's efforts to conform with the special control and any deviations
- Data/information to address issues not covered by guidance documents, special controls, and/or recognized standards
- Information on sterilization, biocompatibility, expiration date, etc., if applicable.

# What to Submit for an Abbreviated 510(k) – FDA Recognized Standards

---

- All the administrative information required for a traditional 510(k)
- An Abbreviated 510(k) that relies on a recognized standard must include a Declaration of Conformity to the Recognized Standard.
- Data/information to address issues not covered by guidance documents, special controls, and/or recognized standards
- Information on sterilization, biocompatibility, expiration date, etc., if applicable.

# When to submit an add-to file

---

- Requesting CLIA categorization after using FDA guidance “Guidance for Industry and FDA Staff; Replacement Reagent and Instrument Family Policy”
- Requesting CLIA categorization for a name change to a cleared device
- Submitting info for CLIA waiver
- Generally not for informing FDA about changes made to your device

## CLIA - Clinical Laboratory Improvement Amendments


[510\(k\)](#) | [Registration & Listing](#) | [Adverse Events](#) | [Recalls](#) | [PMA](#) | [Classification](#) | [Standards](#)
[CFR Title 21](#) | [Radiation-Emitting Products](#) | [X-Ray Assembler](#) | [Medsun Reports](#) | [CLIA](#)

### Search CLIA Database

[Help](#) | [Download Files](#) | [More About CLIA](#)

Test System / Manufacturer

Analyte Name

Show Drop Down ☐


Document Number

Complexity

Analyte Specialty

Effective Date



to



Sort by

Effective Date (descending)



Enter any combination of fields and select Search. You can use the Analyte Drop Down box to select a specific Analyte. For Test System Name/Manufacturer: enter a single word (e.g., Analyzer), an exact phrase (e.g., Acme Analyzer) or multiple words connected by and (e.g., Acme and Analyzer). Hints: Searches are not case sensitive. If you use too many words or misspell a word, you may not find what you are looking for. Consider using Anal to broaden your search to find Analyzer but this will also find Analysis or use full-text search, Select the Go To Simple Search button.

[Go to Simple Search](#)

10



Records per Report Page

[Search](#)
[Clear](#)

# Websites

---

## **Device Advice:**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Overview/default.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Overview/default.htm)

## **How to prepare a special 510(k):**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/PremarketNotification510k/ucm134573.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/PremarketNotification510k/ucm134573.htm)

## **How to prepare an abbreviated 510(k):**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/PremarketNotification510k/ucm134574.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/PremarketNotification510k/ucm134574.htm)

## **FDA standards program:**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm)

# Guidance Documents

---

## **Deciding When to Submit a 510(k) for a Change to an Existing Device (K97-1):**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm080235.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm080235.htm)

## **Frequently Asked Questions on the New 510(k) Paradigm – October 02, 1998**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm073946.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm073946.htm)

# Guidance Documents

---

## **Design Control Guidance For Medical Device Manufacturers**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070627.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm070627.htm)

## **Guidance for Industry and FDA Staff; Replacement Reagent and Instrument Family Policy**

[www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm079185.htm](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm079185.htm)



---

# Thank you!

Doug Rheinheimer | Div of Chemistry and Toxicology Devices | T: (301) 796-6157 |  
F: (301) 847-8513 | [douglas.rheinheimer@fda.hhs.gov](mailto:douglas.rheinheimer@fda.hhs.gov)