

PROPOSED MDR GUIDANCE *CDRH 7-9-13*

Regina J. O'Meara

October 24, 2013


regina.omeara@gmail.com

WHY IS THIS ON THE AGENDA!

- ▶ MDR has been around a long time
- ▶ Most of the audience are IVD Industry
 - ▶ Always been a little grey
 - ▶ Pretty comfortable with FDA expectations
 - ▶ Large firms have robust reporting programs
- ▶ July 9, 2013: New proposed guidance document issued **1st in 16 years**
- ▶ A few significant issues


MDR & IVD DEVICES

“I’m supposed to look at MDRs, but I don’t suspect you have any. I mean what could happen? The instrument falls off the lab bench and breaks the Tech’s foot.” *Ha Ha*




Incorrect or misleading test results impact patient management

MDR & IVD DEVICES



READ IT

- ▶ Available on Website
- ▶ Comments closed
- ▶ Non-binding
- ▶ Q&A format
- ▶ ~ 50 pages
- ▶ Lots of language from 21 CFR Part 803 & 1997 guidance
- ▶ Good list of 3500A mistakes



SIGNIFICANT PROVISIONS

MDR APPLIES TO ALL DEVICES

- ▶ FDAAA 2007 statutory requirement for summary, periodic reports **ONLY** for
 - ▶ Class I & 2
 - ▶ Non-life supporting, non-life sustaining or not permanently implanted
 - ▶ Big deal for IVD
- ▶ No mention of timeline for this rule change
- ▶ FDA is supposed to issue a ruling, and missed this opportunity

HOUSEKEEPING

Annual Certification

Baseline Reporting



DUAL REPORTING BURDEN

- ▶ Foreign Manufacturer + Importer
- ▶ Contract Manufacturer + Spec Developer
 - ▶ Both have to report
 - ▶ Need to obtain FDA exemption for filing
 - ▶ Joint Request

THE 2-YEAR RULE TODAY

- ▶ Once a **MALFUNCTION** has caused or contributed to a serious injury or death
- ▶ Presumption that future occurrences of the malfunction are **LIKELY** to cause or contribute again
- ▶ Thus all recurrences of the malfunction are **REPORTABLE**
- ▶ If after 2 years from the first report there have been **NO** recurrences, the presumption is gone

THE 2-YEAR RULE FUTURE

- ▶ The 2 year rule is gone!
- ▶ Manufacturers have an "**OBLIGATION**" to continue to report all recurrences ...

BUT

- ▶ A manufacturer may submit "**DOCUMENTATION**" that the malfunction has not caused or contributed to additional serious injuries or death, and apply for an exemption

IMPORTANT CLARIFICATIONS

**BE CAREFUL WHAT YOU ASK
FOR**

IMPORTANT CLARIFICATIONS I

- ▶ What does “LIKELY” mean?
 - ▶ Soft Language retained
 - ▶ Catastrophic (to the device) Malfunctions that “may” lead...
 - ▶ Failure of the device that “could” lead...
- ▶ User Error
 - ▶ Incorrect Use if *likely to cause*

IMPORTANT CLARIFICATIONS 2

- ▶ Delay of surgery
 - ▶ No longer need to report based **ONLY** on delay
 - ▶ Delay has to be *likely to cause*
- ▶ Manufacturer's Investigation of MDR Complaints
 - ▶ Good Faith Effort includes a minimum of 1 written request to obtain more info
 - ▶ Firm must analyze even if the device is not returned *w/* similar devices

IMPORTANT CLARIFICATIONS 3

- ▶ Expected device life (record keeping)
 - ▶ **Cannot** use warranty period
 - ▶ Time expected to remain functional including maintenance and calibration cycles
- ▶ Alarms
 - ▶ If an alarm alerts the user before there is any harm, reportable as a malfunction if **likely to cause**

IMPORTANT CLARIFICATIONS 4

- ▶ Risks or Complications in labeling
 - ▶ “Anticipated or intrinsically caused ... not exempt
 - ▶ What does this mean for IVD limitations?
 - ▶ Only time will tell ...

SUMMARY

- ▶ Loss of the 2-year rule increases burden
- ▶ Lack of language supporting FDAAA 2007 is disappointing for IVD Manufacturers
- ▶ Clarification in the Q&A format:
 - ▶ Mixed Blessing
 - ▶ Greater Understanding = Less Interpretation

THANK YOU FOR
YOUR
ATTENTION

