



# A Year in Review: Noteworthy FDA Guidances and Rules for IVD Companies

---

ALLYSON MULLEN

HYMAN, PHELPS & MCNAMARA, P.C.

700 13<sup>TH</sup> STREET NW, SUITE 1200

WASHINGTON, D.C. 20005

# Considerations for Design, Development, and Analytical Validation of NGS

---

FINALS ISSUED

APRIL 13, 2018

- ❖ Significantly modified from the draft.
- ❖ Limited scope going from all NGS-based tests to only those intended to diagnose suspected germline diseases in symptomatic patients.
- ❖ Provides detailed guidance regarding analytical validation, assessing changes, as well as other important topics.
- ❖ Likely de novo, although shortened discussion of the possibility that FDA would classify such tests as 510(k)-exempt

# Use of Public Human Genetic Variant Databases to Support Clinical Validity for Genetic and Genomic-Based *In Vitro* Diagnostics

---

FINALS ISSUED

APRIL 13, 2018

- ❖ Virtually unchanged from the draft guidance, issued July 8, 2016.
- ❖ Describes FDA's considerations in determining whether a publicly available genetic variant database can serve as a source of valid scientific evidence to support the clinical validity of an NGS-based test in a premarket submission.
- ❖ Describes the process by which administrators of such databases can voluntarily apply to FDA for recognition.

# First Formal Recognition

---

- ❖ On December 4, 2018, FDA recognized the first public database containing information about genes, genetic variants, and their relationship to disease.
- ❖ Clinical Genomic Resource (ClinGen) consortium's ClinGen Expert Curated Human Genetic Data.
- ❖ Limited recognition “for germline variants for hereditary disease where there is a high likelihood that the disease or condition will materialize given a deleterious variant (i.e., high penetrance.”

# Consideration of Uncertainty in Making Benefit-Risk Determinations in Medical Device Premarket Approvals, De Novo Classifications, and Humanitarian Device Exemptions

---

DRAFT GUIDANCE

SEPTEMBER 6, 2018

# Factors to Consider:

---

1. Probable benefits;
2. Probable risks;
3. Uncertainty regarding the benefit-risk profile;
4. Patient perspective;
5. Public health need;
6. Feasibility of generating extensive premarket data;
7. Ability to reduce or resolve uncertainty;
8. Likely effectiveness of postmarket mitigations (e.g., labeling);
9. Type of decision being made (e.g., more uncertainty is acceptable for HDEs than PMAs); and
10. Probable benefits of earlier patient access.



# De Novos

---

- ❖ Probable risks will play a large role in analyzing uncertainty
- ❖ Uncertainty of probable benefit can be mitigated if the risks are minimal or through imposition of special controls.

- ❖ Uncertainty connected to amount of clinical data required
  - ❖ Examples include breakthrough devices and devices with small patient populations
  - ❖ IVD-specific example included
- ❖ Correlates the level of uncertainty with the statistical confidence from a submission's clinical study
- ❖ Greater the uncertainty the greater the need for post-market data

# Multiple Function Device Products: Policy and Considerations

---

DRAFT GUIDANCE

APRIL 27, 2018

- ❖ Intended to clarify how FDA assesses the impact of functions that are not subject to FDA review when they are part of a multi-function product that includes at least one function subject to FDA review.
- ❖ FDA will consider:
  - ❖ Whether the other function(s) may impact the safety or effectiveness of the device function-under-review.
  - ❖ Whether there are shared computational resources, data dependencies, or any other type of relationship between the functions.
- ❖ If the other function(s) may impact the safety or effectiveness of the device function-under-review, FDA will consider whether there is an increased risk or adverse effect on performance.
- ❖ Primarily related to software, but could reasonably be applied to IVDs as well

# Benefit-Risk Factors to Consider When Determining Substantial Equivalence in Premarket Notifications (510(k)) with Different Technological Characteristics

---

FINAL GUIDANCE

SEPTEMBER 25, 2018

# 510(k) Review Process

---

- ❖ Do the new device and predicate device have the same intended use?
- ❖ Are there technological differences, and if so, do those differences raise different questions of safety or effectiveness?
- ❖ If no different questions of safety or effectiveness, FDA will then review test methods and data to determine if the data demonstrate substantial equivalence.

The benefit-risk profile of a new device *does not* need to be identical to be as safe and effective as the predicate device.

Risk/Benefit Profile	Assessment Needed?
Increased/Equivalent Benefit Decreased/Equivalent Risk	No
Increased Risk Increase/Equivalent Benefit	Yes
Equivalent Risk Decreased Benefit	Yes
Increased Risk Decreased Benefit	No
Decreased/Equivalent Risk Decreased Benefit	Yes



# Select Updates for Recommendations for Clinical Laboratory Improvement Amendments of 1988 (CLIA) Waiver Applications for Manufacturers of *In Vitro* Diagnostic Devices

## Recommendations for Dual 510(k) and CLIA Waiver by Application Studies

---

DRAFTS RE-ISSUED NOVEMBER 29, 2018

- ❖ The overall requirements and framework of the guidances have not changed substantively since 2017, but virtually all illustrative examples were removed.
- ❖ Leaves open the possibility that a clinical study is not required to establish waived status.
- ❖ CLIA Waiver Guidance: *Four* options for demonstrating waiver, providing additional flexibility in study design
  1. Comparison of a candidate test in the hands of trained and untrained users;
  2. Assay migration study design;
  3. Flex and human factors studies alone; and
  4. Comparison of a candidate test in the hands of untrained users compared to a comparator method in the hands of trained users.

# Voluntary Malfunction Summary Reporting Program for Device Manufacturers

---

FINAL RULE

ISSUED AUGUST 17, 2018

- ❖ Manufacturers of devices and device-led combination products can report certain device malfunctions for low-risk products in summary form on a quarterly basis, as an alternative to Medical Device Reporting requirements.
- ❖ Applies only to reporting of malfunction events by manufacturers
  - ❖ Not deaths or serious injuries, or
  - ❖ Events requiring reporting by importers or user facilities.
- ❖ Only pertains to certain devices whose product code has been in existence for at least two years.

# Summary Malfunction Reporting Schedule

---

<i>Reportable malfunctions or supplemental information that you become aware of during these timeframes:</i>	<i>Must be submitted to FDA by:</i>
January 1 – March 31	April 30
April 1 – June 30	July 31
July 1 – September 30	October 31
October 1 – December 31	January 31

# Additional Limitations

---

- ❖ Does not replace requirement to submit a 5-day report under 21 C.F.R. § 803.53(a).
  - ❖ Including subsequent reportable malfunctions of the same nature involving similar devices
- ❖ Devices that are the subject of a recall
- ❖ New types of reportable malfunctions
- ❖ FDA can revoke eligibility

# De Novo Classification Process

---

PROPOSED RULE

ISSUED DECEMBER 4, 2018

- ❖ Structurally similar to the 510(k) regulation and the PMA regulation
  - ❖ Format and content of a de novo submission,
  - ❖ Procedures governing FDA's review, and
  - ❖ Grounds for denial.



- ❖ Several additional elements of the submission likely to increase the burden on applicants.
  - ❖ Gives FDA authority to inspect manufacturing facilities and clinical trial sites.
  - ❖ Bibliography of all published and unpublished reports on the device and any other information relevant to a device's safety or effectiveness;
  - ❖ Samples of the device and its components, if requested by FDA;
  - ❖ Advertisements for the device.

# Questions?

---

Allyson Mullen

Hyman, Phelps & McNamara, P.C.

700 13<sup>th</sup> Street NW, Suite 1200

Washington, D.C. 20005

[AMullen@hpm.com](mailto:AMullen@hpm.com)